

# Capabilities of industrial computed tomography in the field of dimensional measurements

Horvatić Novak, A.<sup>a,\*</sup>, Runje, B.<sup>a</sup>, Stepanić, J.<sup>a</sup>

<sup>a</sup>University of Zagreb, Faculty of Mechanical Engineering and Naval Architecture, Zagreb, Croatia

## ABSTRACT

The paper discusses the capabilities of industrial computed tomography (CT) in the field of dimensional measurement of products with close tolerances. Computed tomography is a method that allows inspection and measurements of both reachable and unreachable characteristics which makes it very desirable and interesting for application in wide range of industries. In order to evaluate the quality of measurement results obtained by industrial CT, two objects with the same geometry, and made from different materials, were measured. Results obtained with CT were compared with the results obtained by coordinate measuring machine, which were considered to be reference values, and deviations between the results have been analysed. Measurements were repeated five times under repeatability conditions. Repeatability is expressed quantitatively in terms of the dispersion characteristics of the results. Statistical analysis showed that in majority of cases, there were no statistically significant differences between measurement results of equal characteristics obtained at different materials. Obtained deviations in the research could be explained by the fact that the measurements were performed at the industrial CT for general applications. Much better results can be achieved by using a metrology CT device.

© 2017 PEI, University of Maribor. All rights reserved.

## ARTICLE INFO

**Keywords:**  
Metrology  
Dimensional measurement  
Metrological traceability  
Computed tomography (CT)

**\*Corresponding author:**  
[amaliya.horvatic@fsb.hr](mailto:amaliya.horvatic@fsb.hr)  
(Horvatić Novak, A.)

**Article history:**  
Received 27 March 2017  
Revised 15 May 2017  
Accepted 12 June 2017

## References

- [1] Kruth, J.P., Bartscher, M., Carmignato, S., Schmitt, R., De Chiffre, L., Weckenmann, A., (2011). Computed tomography for dimensional metrology, *CIRP Annals – Manufacturing Technology*, Vol. 60, No. 2, 821-842, [doi: 10.1016/j.cirp.2011.05.006](https://doi.org/10.1016/j.cirp.2011.05.006).
- [2] Klobucar, R., Acko, B. (2016). Experimental evaluation of ball bar standard thermal properties by simulating real shop floor conditions, *International Journal of Simulation Modelling*, Vol. 15, No. 3, 511-521, [doi: 10.2507/IJSIMM15\(3\)10.356](https://doi.org/10.2507/IJSIMM15(3)10.356).
- [3] Müller, P., Cantatore, A., Andreassen, J.L., Hiller, J., De Chiffre, L. (2013). Computed tomography as a tool for tolerance verification of industrial parts, *Procedia CIRP*, Vol. 10, 125-132, [doi: 10.1016/j.procir.2013.08.022](https://doi.org/10.1016/j.procir.2013.08.022).
- [4] De Chiffre, L., Carmignato, S., Kruth, J.-P., Schmitt, R., Weckenmann, A. (2014). Industrial application of computed tomography, *CIRP Annals – Manufacturing Technology*, Vol. 63, No. 2, 655-677, [doi: 10.1016/j.cirp.2014.05.011](https://doi.org/10.1016/j.cirp.2014.05.011).
- [5] Horvatić Novak, A., Runje, B. (2016). Dimensional measurement with usage of computed tomography method, In: *Proceedings of 6th International Conference Mechanical Technologies and Structural Materials*, Faculty of Electrical Engineering, Mechanical Engineering and Naval Architecture, Split, Croatia, 51-54.
- [6] Horvatić Novak, A., Runje, B., Butković, D. (2016). Influence of geometrical magnification on computed tomography dimensional measurements, In: *Annals of DAAAM International for 2016, Collection of working papers for 27th DAAAM International Symposium*, DAAAM International Vienna, Vienna, 615-623.
- [7] Welkenhuyzen, F., Kiekens, K., Pierlet, M., Dewulf, W., Bleys, P., Kruth, J.-P., Voet, A. (2009). Industrial computer tomography for dimensional metrology: Overview of influence factors and improvement strategies, In: *Proceedings of the 4th International Conference on Optical Measurement Techniques for Structures and Systems: Optimes2009*, Antwerp, Belgium, 401-410.

- [8] Hiller, J., Reindl, L.M. (2012). A computer simulation platform for the estimation of measurement uncertainties in dimensional X-ray computed tomography, *Measurement*, Vol. 45, No. 8, 2166-2182, doi: [10.1016/j.measurement.2012.05.030](https://doi.org/10.1016/j.measurement.2012.05.030).
- [9] Acko, B., Sluban, B., Tasič, B., Brezovnik, S. (2014). Performance metrics for testing statistical calculations in interlaboratory comparison, *Advances in Production Engineering & Management*, Vol. 9, No. 1, 44-52, doi: [10.14743/apem2014.1.175](https://doi.org/10.14743/apem2014.1.175).
- [10] Müller, P., Hiller, J., Dai, Y., Andreasen, J.L., Hansen, H.N., De Chiffre, L. (2014). Estimation of measurement uncertainties in X-ray computed tomography metrology using the substitution method, *CIRP Journal of Manufacturing Science and Technology*, Vol. 7, No. 3, 222-232, doi: [10.1016/j.cirpj.2014.04.002](https://doi.org/10.1016/j.cirpj.2014.04.002).
- [11] BIMP (2009). JCGM 100:2008 – Evaluation of measurement data – Guide to the expression of uncertainty in measurement, from [http://www.bipm.org/utis/common/documents/jcgm/JCGM\\_100\\_2008\\_E.pdf](http://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf), accessed February 17, 2017.
- [12] BIMP (2009). JCGM 101:2008 – Evaluation of measurement data – Supplement 1 to the “Guide to the expression of uncertainty in measurement” – Propagation of distributions using a Monte Carlo method, from [http://www.bipm.org/utis/common/documents/jcgm/JCGM\\_101\\_2008\\_E.pdf](http://www.bipm.org/utis/common/documents/jcgm/JCGM_101_2008_E.pdf), accessed February 17, 2017.
- [13] ISO 15530-3:2011 (2011). *Geometrical product specifications (GPS) -- Coordinate measuring machines (CMM): Technique for determining the uncertainty of measurement*, ISO copyright office, Geneva, from <https://www.iso.org/standard/53627.html>, accessed March 27, 2017.
- [14] NPL: X-ray computed tomography, from <http://www.npl.co.uk/science-technology/dimensional/x-ray-computed-tomography>, accessed May 7, 2017.
- [15] Acko, B., McCarthy M., Haertig, F., Buchmeister, B. (2012). Standards for testing freeform measurement capability of optical and tactile coordinate measuring machines, *Measurement Science and Technology*, Vol. 23, No. 9, 1-13, doi: [10.1088/0957-0233/23/9/094013](https://doi.org/10.1088/0957-0233/23/9/094013).
- [16] Tasic, T., Acko, B. (2011). Integration of a laser interferometer and a CMM into a measurement system for measuring internal dimensions, *Measurement*, Vol. 44, No. 2, 426-433, doi: [10.1016/j.measurement.2010.11.002](https://doi.org/10.1016/j.measurement.2010.11.002).
- [17] Nikon metrology – CT scanners specifications, from [http://www.nikonmetrology.com/en/EU/Products/X-ray-and-CT-Inspection/Computed-Tomography/XT-H-225-Industrial-CT-Scanning/\(specifications\)](http://www.nikonmetrology.com/en/EU/Products/X-ray-and-CT-Inspection/Computed-Tomography/XT-H-225-Industrial-CT-Scanning/(specifications)), accessed November 8, 2016.
- [18] Schmitt, R., Niggemann, C. (2010). Uncertainty in measurement for x-ray-computed tomography using calibrated work pieces, *Measurement Science and Technology*, Vol. 21, No. 5, doi: [10.1088/0957-0233/21/5/054008](https://doi.org/10.1088/0957-0233/21/5/054008).
- [19] Cantatore, A., Müller, P. (2011). *Introduction to computed tomography*, DTU Mechanical Engineering, Kgs.Lyngby, Denmark.
- [20] Müller, P. (2010). *Use of reference objects for correction of measuring errors in X-ray computed tomography*, DTU Mechanical Engineering, Kgs.Lyngby, Denmark.